Citizen Based Stream Monitoring Program

2009 pH METER CALIBRATION LOG

Oakton Acorn pH Meter Serial Number: _____

Date	Time	Calibration Analyst's Name	Lower pH standard			Upper pH standard			Comments about meter and calibration event.
			pH Buffer	Stabilized pH	Calibrated pH ☺	pH Buffer	Stabilized pH	Calibrated pH ☺	

☺ *If either calibrated pH value is more than 0.2 higher or lower than its corresponding pH buffer value, recalibrate the meter before using the meter in the field. If you need to recalibrate, start a new calibration log entry. Do not edit the entries that you've already recorded.*

Please use one log sheet per each pH Meter. Each monitoring team will record their calibration results on the same log sheet for the monitor they share. Please return CBSM pH meters with their log sheet to: Chris Clayton, 101 S. Webster St. WT/3, PO Box 7921, Madison, WI 53707, (608)266-9260.